

BITREX® QUALITATIVE FIT TEST INSTRUCTIONS



A. EQUIPMENT REQUIRED

1. Hood
2. Nebulizer #1 (BITREX® Sensitivity Solution)
3. Nebulizer #2 (BITREX® Fit Test Solution)
4. Two sets Replacement Nebulizer Inserts
5. BITREX Fit Test Solution Applicator Tube (#0504)
6. BITREX Sensitivity Solution Applicator Tube (#0503)
7. Appropriate Moldex Respirators*

*Use with any Moldex N, R or P class disposable respirator. This fit test kit may also be used with any Moldex half mask respirator by using any N, R or P class filter.

B. NEBULIZER PRE-INSPECTION AND PREPARATION

1. Remove the top half of the nebulizer by unthreading it from the bottom half. (Repeat for each nebulizer)
2. Using your index finger lightly push down on the Atomizer to ensure that it is pushed down all the way onto the bottom half of the nebulizer.
*See Figure 1 (Repeat for each nebulizer).
3. Inspect and ensure that the black O-ring is properly seated. (Repeat for each nebulizer)
4. Replace the top half of nebulizer and remove plug from nebulizer exhaust port. (Repeat for each nebulizer)
5. Use rubber or other protective gloves prior to next step.
6. Get one Sensitivity Solution applicator tube from box of (6) part #0503 and hold upright and squeeze several times until glass ampoule is crushed. Insert applicator tip into the Exhaust Port of the nebulizer labeled "Sensitivity Solution" and squeeze the applicator until drained. Use the foam cushion insert in the Fit Test kit to hold the nebulizer(s) upright. See Figure 2



Figure 1



Figure 2

7. Get one Fit Test Solution applicator tube from box of (6) part #0504 and hold upright and squeeze several times until glass ampoule is crushed. Insert applicator tip into the Exhaust Port of the nebulizer labeled "Test Solution" and squeeze the applicator until drained.
8. Ensure that each nebulizer is able to produce solution mist via the exhaust port by removing plug from Breather Vent and squeeze bulb 4 – 5 times. If solution mist is not visible or seems very weak, carefully (so not to spill the solution) remove the top half of the nebulizer by unthreading it from the bottom half and repeat #2 above. (Repeat for each nebulizer)

C. TASTE SENSITIVITY SCREENING

Purpose: This test is done to ensure that the person being fit tested can detect the taste of BITREX at sufficiently low levels to make the fit test valid.

1. Ensure that the test subject does not eat, drink, smoke or chew gum for 15 minutes before the test.
2. Explain the screening and fit testing procedures to the subject.
3. Instruct subject to put on the hood without a respirator.
4. Position the hood forward so there is about six inches between the subject's face and the window. This is especially important for the fit test. It allows free movement of the head when the subject is wearing a respirator and helps ensure even dispersion of the aerosol around the face seal area.
5. Instruct the test subject to breathe with their mouth slightly open and tongue extended.
6. Using the nebulizer labeled BITREX Sensitivity Solution, inject the aerosol into the hood through the hole in front of the enclosure. Inject ten (10) squeezes of the bulb, fully collapsing and fully expanding the bulb on each squeeze.
7. Ask the subject if he can detect the bitter taste of BITREX.
8. If the subject does not detect bitter taste, inject an additional ten squeezes into the hood.
9. If the subject still does not detect bitter taste, inject an additional ten squeezes, for a total of 30 squeezes.
10. If 30 squeezes were inadequate to elicit a response from the subject, this subject cannot be fit tested with the BITREX test. Another fit test method must be used.
11. If the subject could detect bitter taste, the number of squeezes required to produce a taste response should be noted, i.e., 10, 20 or 30 squeezes, even if subject tasted the BITREX on a number of squeezes other than multiples of ten.
12. Remove the hood. Wait a few minutes, to give the subject time to clear the taste from his mouth, before proceeding to the fit test. A drink of water during this time will aid in removing the bitter taste. Use a paper towel to wipe the subject's mouth of any residue (also wipe mustaches).



BITREX® QUALITATIVE FIT TEST INSTRUCTIONS

D. FIT TEST

1. Use any Moldex disposable N, R or P series class disposable respirator or any Moldex half mask respirator with any N, R or P series filter.
Have the test subject put on and adjust the respirator per the instructions provided with the respirator. Have the subject select the size respirator that provides the best fit. The subject may find a mirror useful in the adjustment process. The subject should wear the respirator at least five (5) minutes before starting the test.
2. Instruct the subject to put on and position the hood as before and breathe with their mouth slightly open and tongue extended.
3. Using the nebulizer labeled BITREX Fit Test Solution, inject the fit test aerosol into the hood. The same number of squeezes is required as was necessary to elicit a response in the threshold sensitivity screening test, i.e., 10, 20 or 30 squeezes.
4. To maintain an adequate concentration of aerosol during the test, one half of the initial number of squeezes should be injected again every 30 seconds (i.e., 5, 10, or 15) throughout the test.
5. After injecting the aerosol initially, ask the test subject to perform the following exercises for 60 seconds each:
 - (a) Normal breathing.
 - (b) Deep breathing.
 - (c) Turning head from side to side, stopping at each end of travel for one or two breaths.
 - (d) Moving head up and down, stopping at each end of travel for one or two breaths.
 - (e) Talking, reciting the alphabet, or reading a prepared text (the Rainbow Passage is recommended, see below).
 - (f) Jogging in place.
 - (g) Normal breathing.
6. Instruct the subject to indicate if they detect the bitter taste at any time during the test.
7. If the entire test is completed without the subject detecting the bitter taste of the aerosol, the test is passed and the respirator's fit on that individual is judged adequate.
8. At any time during the test, if the subject detects the bitter taste of the aerosol the test is stopped at this point. When this occurs, the fit of the respirator on the subject is judged inadequate.
9. If the test is failed, before retesting the subject, a 15-minute waiting period must be observed and the taste sensitivity screening test must be performed again. Test again with same Moldex model.
10. If the second test fails, repeat tests with another size and/or model Moldex respirator.
11. Document results of fit tests for each employee.

OSHA

This fit test meets the performance criteria for fit testing half mask respirators under current OSHA Standards.

OSHA regulation 1910.134 (f) requires that the user be fit tested before being assigned a respirator.

CLEANING

Immediately after completing the test, discard any unused solutions. Rinse the nebulizers with warm water to prevent clogging. Wipe inside of the hood with a damp cloth or paper towel to remove any deposited Test Solution.

FOR TECHNICAL ASSISTANCE CALL MOLDEX TECHNICAL SERVICE DEPARTMENT

800/421-0668 ext. 512/550, or 310/837-6500 ext. 512/550.

LIMITED WARRANTY IMPORTANT NOTICE TO PURCHASER

This limited warranty is made in lieu of the warranties of merchantability, fitness for particular purposes and all other warranties, express or implied. There are no other warranties which extend beyond the description on the face hereof. The physical standards and specifications of Moldex will be met by products sold. **Exclusive Remedies:** damages for the breach of this limited warranty are limited to the replacement of such quantity of Moldex, products proved to be defectively manufactured. Except as provided above, Moldex shall not be liable or responsible for any loss, damage, or liability, direct, indirect, incidental, special, or consequential, arising out of sale, use, or misuse, or the inability to use products by the user.

DISTRIBUTED BY:

RAINBOW PASSAGE

When the sunlight strikes raindrops in the air, they act like a prism and form a rainbow. The rainbow is a division of white light into many beautiful colors. These take the shape of a long arch, with its path high above, and its two ends apparently beyond the horizon. There is, according to legend, a boiling pot of gold at one end. People look, but no one ever finds it. When a man looks for something beyond his reach, his friends say he is looking for the pot of gold at the end of the rainbow.

MOLDEX-METRIC, INC.

10111 W. Jefferson Blvd. Culver City, CA 90232
800/421-0668 or 310/837-6500 FAX 310/837-9563
www.moldex.com e-mail: sales@moldex.com

CANADIAN CUSTOMER SERVICE

TEL: 310/837-6500, Ext. 517
Fax: 310/837-9563
Made in the U.S.A. 9700-702 REV H 07/08

Moldex is a registered trademark and the mesh design is a trademark of Moldex-Metric, Inc.
U.S. Patents #D285,374 and #4,850,374



MOLDEX
Ideas that wear well.