

IV. MOLDEX INDUSTRIAL 42 CFR 84 DISPOSABLE RESPIRATOR USER INSTRUCTIONS

Each specific model may contain additional warnings and instructions.
You **MUST** refer to the packaging for specific instructions.

A. APPLICATIONS

1. WARNING TO USER

- a. Follow all instructions and warnings on the use of this respirator and wear during all times of exposure. Failure to do so will reduce respirator effectiveness, wearer protection and may result in sickness or death.

- b. Before use, the user must first be trained by the employer in proper respirator use, in accordance with applicable safety and health standards, for the contaminant and exposure level in the assigned work area.
- c. The particulates and other contaminants which can be dangerous to your health include those which you cannot see, taste, or smell.
- d. This product has not been sold with warnings or use instructions for personnel involved in healthcare or related situations, where there may be the possibility of contact with disease or biological hazards. If you are considering such uses, first call the Moldex Technical Service Dept., +1 (310) 837-6500 or +1 (800) 421-0668 ext. 512/550.

USE AGAINST (N95, N99 & N100 MODELS):

1. Particulate aerosols free of oil.
2. For 2400, 2800/1, 4800 only, nuisance levels (less than OSHA Permissible Exposure Limits) of Organic Vapor Odors and/or Ozone.
3. For 2500 only, nuisance levels (less than OSHA Permissible Exposure Limits) of Acid Gas Irritants.

USE AGAINST (R95, P95 & P100 MODELS):

1. Oil and non-oil particulate aerosols.
2. For 2840/1R95 only, nuisance levels (less than OSHA Permissible Exposure Limits) of Organic Vapor Odors and/or Ozone.
3. For 2940/1R95 only, nuisance levels (less than OSHA Permissible Exposure Limits) of Acid Gas Irritants.

DO NOT USE AGAINST (N95, N99 & N100 MODELS):

1. Concentrations of contaminants which are unknown or are immediately dangerous to life or health.
2. Concentrations of particulates which exceed the maximum use concentration or 10 times the OSHA Permissible Exposure Limit, whichever is lower.
3. Gasses, vapors, asbestos, paint spray, sandblasting or particulate materials which generate harmful vapors. This respirator is not approved for any vapors or gases, so when used against Organic Vapor Odors and/or Ozone or Acid Gas Irritants do not exceed the OSHA Permissible Exposure Limit.
4. Oil-Based mists

DO NOT USE AGAINST (R95, P95 & P100 MODELS):

1. Concentrations of contaminants which are unknown or are immediately dangerous to life or health.
2. Concentrations of particulates which exceed the maximum use concentration or 10 times the OSHA Permissible Exposure Limit, whichever is lower.

3. Gases, vapors, asbestos, paint spray, sandblasting or particulate materials which generate harmful vapors. These respirators are NOT approved for any vapors or gases, so when used against Organic Vapor Odors and/or Ozone or Acid Gas Irritants, do not exceed the OSHA Permissible Exposure Limit.
4. Oil-based mists for longer than one 8 hour work shift.

RESTRICTIONS (N95, N99 & N100 MODELS):

1. This respirator does not supply oxygen and must not be used in atmospheres containing less than 19.5% oxygen.
2. Do not remain in contaminated area if any physical distress occurs, for example breathing difficulty, dizziness or nausea.
3. Leave contaminated area and replace respirator if it is damaged, distorted, a proper fit can not be obtained or breathing becomes difficult.
4. Prior to each use, carefully inspect the entire respirator, including filter media and strap/weld attachment area for tears and damage. Staple or strap/weld attachment perforations do not affect respirator performance.
5. Do not alter, modify, or abuse this respirator.
6. Store unused respirators in box/bag in a clean, dry, non-contaminated area.
7. Dispose of respirator according to your employer's policy and local regulations.
8. Dispose no later than thirty days after first use.
9. Use respirator before the "use by" expiration date printed on box/bag.
10. If used for welding, wear appropriate eye and face protection.

RESTRICTIONS (R95, P95 & P100 MODELS):

1. This respirator does not supply oxygen and must not be used in atmospheres containing less than 19.5% oxygen.
2. Do not remain in contaminated area if any physical distress occurs, for example breathing difficulty, dizziness or nausea.
3. Leave contaminated area and replace respirator if it is damaged, distorted, a proper fit can not be obtained or breathing becomes difficult.
4. Prior to each use, carefully inspect the entire respirator, including filter media and strap/weld attachment area for tears and damage. Staple or strap/weld attachment perforations do not affect respirator performance.
5. Do not alter, modify, or abuse this respirator.
6. Store unused respirators in box/bag in a clean, dry, non-contaminated area.
7. Dispose of respirator according to your employer's policy and local regulations.
8. Where oil mists are present dispose no later than 8 hours after first use. Where non-oil particulates only are present, dispose no later than 30 days after first use.
9. Use respirator before the "use by" expiration date printed on box/bag.
10. If used for welding, wear appropriate eye and face protection.

B. APPROVALS

1. Review NIOSH/DHHS label in approvals section of disposable respirator box or bag.

MOLDEX HEALTHCARE PARTICULATE RESPIRATORS/ SURGICAL MASK USER INSTRUCTIONS

Each specific model may contain additional warnings and instructions.
Refer to the packaging for specific instructions.

A. APPLICATIONS

1. WARNING TO USER

- a. Non-latex straps are used in the headbands of these respirators. However, individuals highly sensitive to natural rubber latex may have an allergic reaction.
- b. This product does not eliminate the risk of contracting any disease or infection. Change immediately if the respirator comes in contact with blood or fluids, including body fluids.
- c. OSHA has not set a permissible exposure level for airborne biohazards.
- d. This product has not been sold with warning or use instructions for personnel involved in industrial or related situations. If you are considering such uses, first contact the Moldex Technical Services Department at +1 (800) 421-0668 or +1 (310) 837-6500, ext. 512/550.

2. RESTRICTIONS

- a. Before use, a written respiratory protection program in accordance with 29 CFR 1910.134 must be implemented.
- b. This respirator does not supply oxygen and must not be used in atmospheres containing less than 19.5% oxygen.
- c. Prior to each use, carefully inspect the entire respirator, including filter media and strap attachment area for tears and damage. Staple or strap attachment perforations do not effect NIOSH approval.
- d. If respirator is damaged, distorted, a proper fit cannot be obtained or breathing becomes difficult, leave contaminated area as soon as possible and replace respirator.
- e. If the respirator comes in contact with blood or fluids, including body fluids, leave contaminated area as soon as possible and discard and replace the respirator.
- f. Do not alter, modify or abuse this respirator.
- g. Store unused respirators in box/bag in a clean, dry, non-contaminated area.
- h. Dispose of respirator according to your employer's policy and local regulations.

- i. Do not reuse or store for reuse or hang around neck unless your employer specifically authorizes reuse.
- j. When used for surgical procedures, discard after every use.
- k. Use respirator before the “use by” expiration date printed on box/bag.

For technical assistance call Moldex Technical Service Department, +1 (800) 421-0668 or +1 (310) 837-6500, ext. 512/550.

3. INTENDED USE

The various models of Moldex Type N95 Healthcare Particulate Respirators and Surgical Masks meet CDC Guidelines for TB Exposure Control within healthcare facilities. These devices are also intended to be worn by healthcare personnel during surgical procedures to protect both the patient and healthcare personnel from the transfer of microorganisms, body fluids and particulate material.

4. CONTRAINDICATIONS

- a. Not for use with beards, or other facial hair that prevents direct contact between the face and sealing surface of the respirator.
- b. Eyewear must not prevent direct contact between the face and sealing surface of the respirator.
- c. Not to be used on children.

5. DESCRIPTION

The Moldex Healthcare Series N95 Particulate Respirator and Surgical Mask is designed to help provide respiratory protection for the wearer. This product has been tested¹ and certified by NIOSH as an N95 respirator and as having a filter efficiency level of 95% or greater against particulate aerosols free of oil. It is fluid resistant², disposable and may be worn in surgery or throughout the hospital.

¹ Tested in accordance with NIOSH 42 CFR 84.

² Passed ASTM F 1862 @ 160 mm Hg.

B. APPROVALS

1. Review NIOSH/DHHS label in approvals section of disposable respirator box or bag.



DEPARTMENT OF HEALTH & HUMAN SERVICES
NATIONAL INSTITUTE FOR OCCUPATIONAL SAFETY AND HEALTH

NOVEL LLC
CULVER CITY, CALIFORNIA, U.S.A.
866-546-6835



THESE RESPIRATORS ARE APPROVED ONLY IN THE FOLLOWING CONFIGURATIONS:

TC- PROTECTION ¹	RESPIRATOR	RESPIRATOR					CAUTIONS AND LIMITATIONS ²
		1510	1511	1512	1513	1517	
84A-4333	N95	X					A B C J M N O P
84A-0013	N95		X	X	X	X	A B C J M N O P

1. Protection

N95-Particulate Filter (95% filter efficiency level) effective against particulate aerosols free of oil; time use restrictions may apply

2. Cautions and Limitations

A- Not for use in atmospheres containing less than 19.5 percent oxygen.
 B- Not for use in atmospheres immediately dangerous to life or health.
 C- Do not exceed maximum use concentrations established by regulatory standards.
 D- Failure to properly use and maintain this product could result in injury or death.
 M- All approved respirators shall be selected, fitted, used, and maintained in accordance with MSHA, OSHA, and other applicable regulations.
 N- Never substitute, modify, add, or omit parts. Use only exact replacement parts in the configuration as specified by the manufacturer.
 O- Refer to User's instructions, and/or maintenance manuals for information on use and maintenance of these respirators.
 P- NIOSH does not evaluate respirators for use as surgical masks.

SAMPLE LABEL